

# United States Patent and Trademark Office

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	N NO. FILING DATE		FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.	
10/716,352	11/18/2003		Shazhou Zou	7629		
75	90	06/20/2005		EXAMINER		
Tianxin Wang 9768 Early Spring Way				BOCKELMAN, MARK		
Columbia, MD			ART UNIT	PAPER NUMBER		
				3762		

DATE MAILED: 06/20/2005

Please find below and/or attached an Office communication concerning this application or proceeding.

				SA				
,		Application No.	Applicant(s)					
Office Action Summary		10/716,352	ZOU ET AL.					
		Examiner	Art Unit					
		Mark W. Bockelman	3762					
The N Period for Repl	IAILING DATE of this communication ap y	pears on the cover sheet with t	the correspondence addre	ss				
THE MAILIN  - Extensions of ti after SIX (6) Mo  - If the period for  - If NO period for  - Failure to reply Any reply receive	IED STATUTORY PERIOD FOR REPL G DATE OF THIS COMMUNICATION. me may be available under the provisions of 37 CFR 1. DNTHS from the mailing date of this communication. reply specified above is less than thirty (30) days, a repreply is specified above, the maximum statutory period within the set or extended period for reply will, by statutived by the Office later than three months after the mailing erm adjustment. See 37 CFR 1.704(b).	136(a). In no event, however, may a reply bly within the statutory minimum of thirty (30 I will apply and will expire SIX (6) MONTHS te, cause the application to become ABAND	be timely filed  0) days will be considered timely.  6 from the mailing date of this common DONED (35 U.S.C. § 133).	unication.				
Status								
1)⊠ Respo	nsive to communication(s) filed on 14/1	<u> March 2005</u> .						
2a)⊠ This ad	n)⊠ This action is <b>FINAL</b> . 2b)□ This action is non-final.							
3)☐ Since t	Since this application is in condition for allowance except for formal matters, prosecution as to the merits is							
closed	in accordance with the practice under	Ex parte Quayle, 1935 C.D. 1	1, 453 O.G. 213.					
Disposition of C	Claims							
	s) <u>1-18</u> is/are pending in the application							
4a) Of	the above claim(s) is/are withdra	awn from consideration.						
· <u></u>	s) is/are allowed.							
	s) <u>1-18</u> is/are rejected.			'				
	s) is/are objected to.							
8) Claim(	s) are subject to restriction and/	or election requirement.						
Application Pap	pers							
9)∐ The sp	ecification is objected to by the Examin	er.						
10)∏ The dra	☐ The drawing(s) filed on is/are: a)☐ accepted or b)☐ objected to by the Examiner.							
• •	Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).							
	Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).							
11)∐ The oa	th or declaration is objected to by the E	Examiner. Note the attached O	ffice Action or form PTO-	152.				
Priority under 3	5 U.S.C. § 119							
a)	viedgment is made of a claim for foreign b) Some * c) None of:  Certified copies of the priority document Certified copies of the priority document Copies of the certified copies of the priority document Copies of the certified copies of the priority document Copies of the certified copies of the priority document Copies of the certified copies of the priority document Copies of the certified copies of the priority document Copies of the certified copies of the priority document Copies of the certified copies of the priority document Copies of the certified copies of the priority document Copies of the certified copies of the priority document Copies of the certified copies of the priority document Copies of the priority doc	nts have been received. Its have been received in Apploity documents have been recaute (PCT Rule 17.2(a)).	lication No ceived in this National Sta	nge				
* See the	attached detailed Office action for a lis	it of the certified copies not rec	eived.					
Attachment(s)								
	rences Cited (PTO-892) tsperson's Patent Drawing Review (PTO-948)		nmary (PTO-413) fail Date					
	sclosure Statement(s) (PTO-1449 or PTO/SB/08	——————————————————————————————————————	mal Patent Application (PTO-15	2)				

#### **DETAILED ACTION**

# Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

- (b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.
- (e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

Claims 1-6, 8-12 are rejected under 35 U.S.C. 102(b) as being anticipated by Bertolucci USPN 4,981,146.Bertolucci shows a small housing36, one electrode assembly (20 and or 28) that can be linked to the housing by wire or band (electrode 20) or some other means conductive means (adhesive or solder or screw- for electrode 28), circuit means (figure 3), and band member 42. The circuitry sends pulsed stimulation signals (see column 5 lines 1-10) which may be varied in amplitude (column 4 lines 20 –25) and includes an on/off switch (column 4 lines 29-31). The device is used in treating nausea. The electrode 20 is external of the housing as well as the fastening means.

## Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and

Application/Control Number: 10/716,352

Art Unit: 3762

the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

Claims 1-8, 10-12 rejected under 35 U.S.C. 102(e) as anticipated by or, in the alternative, under 35 U.S.C. 103(a) as obvious over Michelson et al. USPN 6,445,955.

In figure 13 and 13a Michelson shows a housing 30 that is worn on the wrist containing circuitry and what appears to be a attachment band. The device communicates with element 1 which is linked to the housing and has electrodes 5 for stimulation. The device has an acupuncture stimulation mode (column 2 lines 57-62) with a frequency of about 100 Hz and is adjustable (column 11 lines 10-40). The examiner considers the "worn on the wrist" comment regarding figure 13 to be a teaching of a means for attaching the device or otherwise it would have been obvious to have included elements such as a band to allow it to be worn about the wrist.

Claims 13-18 are rejected under 35 U.S.C. 103(a) as being unpatentable over Bertolucci USPN 4,981,146 or Michelson et al. USPN 6,445,955. While both of the references fail to discuss any of the specific acupuncture points that applicant recites in the method claims. Bertolucci at least teaches the treatment of nausea and would likely use the same site and regardless, to have used the Bertolucci and Michelson devices on whatever acupuncture point desired would have been obvious.

Claim 9 is rejected under 35 U.S.C. 103(a) as being unpatentable over Michelson et al. USPN 6,445,955 in view of Bertolucci USPN 4,981,146. Although no on/off switch

is found in Michelson to have include such would have been conventional as evidenced by Bertolucci.

# Response to Arguments

Applicant's arguments filed 3-14-2005 have been fully considered but they are not persuasive. Contrary to applicant's arguments the linkable electrodes 20, 28 in Bertocucci are considered external to the fastening means and on the housing surface. Thus the electrode is considered to be external. With regard to Michelson et al, applicant has not addressed the embodiment described in figure 13, which examiner relies upon.

### Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of

Art Unit: 3762

the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Mark W Bockelman whose telephone number is (571) 272-4941. The examiner can normally be reached on Monday - Friday 8:00 to 4:30.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Angela Sykes can be reached on (571) 272 -4955. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

**MWB** 

June 12, 2005